



Madhya Pradesh State Electronics Development Corporation Ltd.

State IT Centre, 47-A, Arera Hills, Bhopal – 462 011

CORRIGENDUM-1

Date: 18th January, 2016

Tender No. MPSEDC/ETDC/2015/301 DATED 29.12.2015

Supply and Installation of Calibrators for Testing and Calibration of Equipments as per BIS Standard at MPSEDC Calibration Lab, Bhopal (M.P.)

*The last date of submission of offers has been extended from 20th January, 2016 to **27th January 2016 (Wednesday)** at 3:00 pm. The technical bid will be opened on the same day at 4:00 pm.*

1. CREEPAGE AND CLEARANCE TEST SET FOR TESTS CONFIRMING TO IEC 60950-1: 2001, IEC 60950-1:2005, IEC 60065:2005 SAFETY REQUIREMENTS

	ORIGINAL SPECIFICATION AS PER TENDER		AMENDED SPECIFICATIONS	
Sl. No.	PARAMETER	SPECIFICATION	PARAMETER	SPECIFICATION
1	--		Creepage and Clearance Set	
	Stainless Steel 0.5mm to 10mm	32 gauges from 0.5 to 10mm	Stainless Steel 0.5mm to 10mm	32 gauges from 0.5 to 10mm
		Accuracy 0.1mm from 0.5 to 2mm		Accuracy 0.1mm from 0.5 to 2mm
		Accuracy 0.5mm from 2 to 10mm		Accuracy 0.5mm from 2 to 10mm
2	Digital Micrometer		Digital Micrometer	
	Size/ Range:	0-1 in./ 0-25mm.	Size/ Range:	0-25 mm.
	Resolution:	0.0005 in./ 0.001m	Resolution:	0.0001mm
	Accuracy:	0.00005in./ 0.001mm	Accuracy:	±5µm
3	Digital Vernier Caliper		Digital Vernier Caliper	
	Size/ Range:	0-6 in./0-150mm	Size/ Range:	0-300mm
	Resolution:	0.0005 in./0.01m	Resolution:	0.01mm
	Accuracy:	0.001in./0.03mm	Accuracy:	0.03mm
4	Optical Comparator Gauge Set		Optical Comparator Gauge Set	
	Magnification	10X	Magnification	10X
	Field of view	25mm(1 inch)	Field of view	25mm(1 inch)
	Working Distance	50mm(2 inch) Approx.	Working Distance	50mm(2 inch) Approx.
	Reticle	Glass.	Reticle	Glass.
5	Accessories		Accessories	
6	NABL Calibration Certificate with Data.		NABL Calibration Certificate with Data.	



Madhya Pradesh State Electronics Development Corporation Ltd.

State IT Centre, 47-A, Arera Hills, Bhopal – 462 011

7	Users Manual	Users Manual
8	Warranty Minimum One Year.	Warranty Minimum One Year.

All other terms & conditions remain unchanged

sd/-

Chief General Manager

M P State Electronics Dev. Corp. Ltd.

State IT Centre, 47-A, Arera Hills Bhopal – 462011,

Phone: 0755-2518754, 2518603 Fax: 0755-2579824